Docusign Envelope ID: A5FD9A2D-6B61-8CCF-82D3-F38D22B7B2D2

00030842 Rev Q

Declaration of Conformity

Responsible Manufacturer: Laerdal Medical AS
Seehusens gate 1, P.O. Box 377
4002 Stavanger, Norway

Manufacturing sites:
ShockLink and Remote Control:
Eastek International (Malaysia) Sdn. Bhd., 65 Barrage Industrial Estate, Mergong, 05150 Alor Setar, Kedah, Malaysia
Schiller ShockLink Training Adapter, Corpuls Training Adapter, Tempus ALS ShockLink Training Adapter, Zoll ShockLink Training Adapter:
Cablex Wire & Cable (Kunshan) MFG, No. 168, Binjiang South Road, Zhangpu Town, Kunshan, Suzhou City, Jiangsu Province, China
Physio/Mindray ShockLink Training Adapter:
Amphenol Alden, No.7 Dongbao Road, Dongkeng community, Fenghua sub district, Guangming district, 518107 Shenzhen, China

Product Name:	**ShockLink**
Product Options:	
185-10050	ShockLink system
185-20050	ShockLink only
Accessories:	
185-50050	Physio/Mindray ShockLink Training Adapter
185-50250	Schiller ShockLink Training Adapter
185-50550	Tempus ALS ShockLink Training Adapter
185-50350	Corpuls Training Adapter
185-50450	Zoll ShockLink Training Adapter
185-50750	Nihon Kohden ShockLink training adapter
185-00350	Remote Control ShockLink

to which this declaration relates is in conformity with **EU Council Directive 2014/53/EU on Radio Equipment (RED)** and **Council Directive 2011/65/EU on Restriction on the use of certain hazardous substance (RoHS)**, as amended
And **UK Electronic Compatibility Regulation SI 2016/1091** and **RoHS Regulation SI 2012/3032**

All supporting documentation is retained by the manufacturer.

The conformity is based on the following standards:
EMC:
EN 62479:2010
EN 61000-6-1:2007
EN 61000-6-3:2007/A1:2011
Safety:
EN 62368-1:2014 + A11:2017
Radio:
EN 301 489-1 V1.9.2
EN 301 489-17 V2.2.1
ETSI EN 300 328 V2.2.2

DocuSigned by:
Mari Kaada
SIGNATURE 5AD2CD13D8AA495...

Corporate Director Q&R
POSITION IN ORGANISATION

Mari Kaada
PRINT NAME

2026-08-27
DATE